

شهادة اعتماد

رقم ML 008



يقر نظام الاعتماد العراقي بأن:

مختبر الكواشف الكيميائية للتحليلات الطبية

العراق - بغداد - الكاظميه - مدينه الهبنة - شارع ٦٠ - قرب مستشفى الكاظميه التعليمي

تم اعتماده وفقا لمتطلبات المواصفة ISO 15189:2012
(المختبرات الطبية - متطلبات الكفاءة والجودة)

في مجال:

- اختبارات الكيمياء السريرية

شرط التوافق مع متطلبات المواصفة اعلاه ومتطلبات IQAS الخاصة بالاعتماد
مجال الاعتماد المرفق بالشهادة يعتبر جزءا لا يتجزء منها
يمكن الحصول على الاصدار الاحدث من مجال الاعتماد من خلال الموقع الالكتروني

<https://iqas.mop.gov.iq>

يكون الاعتماد نافذا من ٢٠٢٣/٧/٢٥ الى ٢٠٢٥/٧/٢٤

تاريخ منح الاعتماد لأول مرة

٢٠٢١/٤/٢٩

شتم

أ.د. محمد علي تميم
نائب رئيس مجلس الوزراء
وزير التخطيط

٢٠٢٢ / ٨ / ٢١

د. محمد لطيف أحمد
مدير عام الهيئة العراقية للاعتماد/ وكالة

Ministry of planning
Iraqi Organization for Accreditation
IQAS

ACCREDITATION CERTIFICATE

No. ML 008



Iraqi Accreditation System Certify that:

Al-Kawashaf Chemical Laboratory

Iraq – Baghdad - Al-Kadhimiya –Al -hbnah city- Street 60 -near Al-Kadhimiya teaching hospital

Is accredited according to the requirements of the standard ISO 15189:2012
(Medical laboratories — Requirements for quality and competence)

In the field of:


- Clinical Chemistry Testing

This accreditation is subject to with the above standard & IQAS requirements
The scope of accreditation is attached to the certificate & considered as part of it


The most recent issue of the accreditation scope is available on IQAS website
<https://iqas.mop.gov.iq>

Accreditation is valid From 25/7/2023 To 24/7/2025

Initial accreditation date
29/4/2021


Dr. Mohammed Lateef Ahmed
General Manager of IQAS


Dr. Mohammed Ali Tamim
Deputy Prime Minister
Minister of Planning
21-8-2023

	استمارة مجال الاعتماد Scope of Accreditation form	نظام الاعتماد العراقي IQAS
Organization address: Iraq – Baghdad- -Al- Kadhimiya-Al- Al hbnah city- Street 60-near Al-Kadhimiya Teaching Hospital	Organization name : Al-Kawashaf Chemical Laboratory	Accreditation no.: ML 008
Signature:  Abdul Wahid Mohammed Ibrahim Deputy General Manager	Accreditation is valid: From 25/7/2023 To 24/7/2025	Issue no. : 002

Branch/ Sector	Parameters/ Test	Type of Sample or System Material	Principle of Measurement	Instrument	Units
Clinical chemistry	Glucose	Serum	Hexokinase	Randox RX imola	mg/dl
	Cholesterol		Cholesterol Oxidase - Abell Kendall		mg/dl
	Triglyceride		Lipase/GPO-PAP no correction		mg/dl
	Creatinine		Enzymatic UV method (340nm)		mg/dl
	Urea		Urease, kinetic		mg/dl
	Calcium		Arsenazo		mg/dl
	Phosphate inorganic		Phosphomolybdate UV		mg/dl
	ALT		Tris buffer without P5P		U/l
	AST		Tris buffer without P5P		U/l
	Uric Acid		Uricase Perox. with ascorb. ox @ 546nm		mg/dl
	Albumin		Bromocresol Green		g/dl