

شهادة اعتماد

رقم 014 CL



يقر نظام الاعتماد العراقي بأن:
مختبر شركة النبأ اليقين للخدمات والاستشارات الهندسية
العراق - اقليم كردستان - اربيل

تم اعتماده وفقا لمتطلبات المواصفة ISO/IEC 17025:2017
(المتطلبات العامة لاهلية مختبرات الفحص والمعايرة)

- في مجال:
- معايرة درجة الحرارة
 - معايرة الضغط
 - معايرة الكهرباء

شرط التوافق مع متطلبات المواصفة اعلاه ومتطلبات IQAS الخاصة بالاعتماد
مجال الاعتماد المرفق بالشهادة يعتبر جزءا لا يتجزأ منها

يمكن الحصول على الاصدار الاحدث من مجال الاعتماد من خلال الموقع الالكتروني

<https://iqas.mop.gov.iq>

يكون الاعتماد نافذا من 2025/7/30 الى 2027/7/29
تاريخ منح الاعتماد لأول مرة
2025/7/30

لقد
أ.د. محمد علي تميم
نائب رئيس مجلس الوزراء
وزير التخطيط

المهندس
عبد الواحد محمد ابراهيم
مدير عام الهيئة/ وكالة

Ministry of planning
Iraqi Organization for Accreditation
IQAS

ACCREDITATION CERTIFICATE

No. CL 014



Iraqi Accreditation System Certify that:

Laboratory of Al-Nabaa Al-Yaqeen Co. for Engineering Service and Consultancies

Iraq-Kurdistan region- Erbil

Is accredited according to the requirements of the standard ISO/IEC 17025:2017
(General Requirements for the Competence of Testing and Calibration Laboratories)

In the field of:

- Temperature Calibration
- Pressure Calibration
- Electricity Calibration

This accreditation is subject to with the above standard & IQAS requirements
The scope of accreditation is attached to the certificate & considered as part of it


The most recent issue of the accreditation scope is available on IQAS website
<https://iqas.mop.gov.iq>

Accreditation is valid From 30/7/2025 To 29/7/2027

Initial accreditation date
15/7/2021


Eng. Abdul Wahid M. Ibrahim
Director General of IQAS


Dr. Mohammed Ali Tamim
Deputy Prime Minister
Minister of Planning

	استمارة مجال الاعتماد Scope of Accreditation form	نظام الاعتماد العراقي IQAS
Organization address: Iraq-Kurdistan region- Erbil	Organization name: Laboratory of Al-Nabaa Al-Yaqeen Co. for Engineering Service and Consultancies	Accreditation no.: CL 014
Signature:  Eng. Abdul Wahid M. Ibrahim Director General of IQAS	Accreditation is valid: From 30/7/2025 To 29/7/2027	Issue no.: 003

Calibration Field	Type of Instrument or Material to be Calibrated or Measured	Type of Instrument or Material to be Calibrated or Measured	Calibration Method	Measurement Range	Uncertainty ($\pm$)	Location
Pressure	Pressure gauge	Using Reference Pressure Gauge fluke p 3124/3125 deadweight tester	CP/M-01,03 DKD-R 6-1:2014	0-300 bar	0.2 bar	laboratory
Pressure	Pressure gauge	Using Reference Pressure Gauge fluke p 3124/3125 deadweight tester	CP/M-01,03 DKD-R 6-1:2014	300-1100 bar	0.4 bar	laboratory
temperature	dial temperature indicator with sensor	dry block TA-650P SN:1680617	EURAMET cg -8	50 to 650 °C	2°C	laboratory
temperature	dial temperature indicator with sensor	TC 150-TEK-KNOW	EURAMET cg -8	-25 to 150 °C	2°C	laboratory
Electrical	voltage measurement AC-DC	portable multi-function calibrator SN:V01586H15	EURAMET cg -15	0 – 1000 V	0.003 V	laboratory
Electrical	Current measurement AC-DC	portable multi-function calibrator SN:V01586H15	EURAMET cg -15	0 -500 A	0.003 A	laboratory
Electrical	Resistance measurement	portable multi-function calibrator SN:V01586H15	EURAMET cg -15	0 -10 M Ω	0.16 M Ω	laboratory
Electrical	thermocouple output	portable multi-function calibrator SN:V01586H15	EURAMET cg -15	0 -1200 °C	0.16 °C	laboratory