

شهادة اعتماد

رقم CL 016



يقر نظام الاعتماد العراقي بأن:
مختبرات شركة بروفشنال لابروتوري للمقاولات العامة/ المحدودة
العراق- اقليم كردستان- اربيل

تم اعتمادها وفقا لمتطلبات المواصفة ISO/ IEC 17025:2017
(المتطلبات العامة لاهلية مختبرات الفحص والمعايرة)

- في مجال:
- معايرة القوة
 - معايرة الكتلة
 - معايرة درجة الحرارة

شرط التوافق مع متطلبات المواصفة اعلاه ومتطلبات IQAS الخاصة بالاعتماد
مجال الاعتماد المرفق بالشهادة يعتبر جزءا لا يتجزأ منها

يمكن الحصول على الاصدار الاحدث من مجال الاعتماد من خلال الموقع الالكتروني
<https://iqas.mop.gov.iq>

يكون الاعتماد نافذا من 2025/7/10 الى 2027/7/9
تاريخ منح الاعتماد لاول مرة
2025/7/10

أ.د. محمد علي تميم
نائب رئيس مجلس الوزراء
وزير التخطيط

المهندس
عبد الواحد محمد ابراهيم
مدير عام الهيئة/ وكالة

Ministry of planning
Iraqi Organization for Accreditation
IQAS

ACCREDITATION CERTIFICATE

No. CL 016



Iraqi Accreditation System Certify that:

Professional Laboratory Company for General Contracting Ltd

Iraq - Kurdistan Region - Erbil

Is accredited according to the requirements of the standard ISO/IEC17025:2017
(General Requirements for the Competence of Testing and Calibration Laboratories)

In the field of:

- Force Calibration
- Mass Calibration
- Temperature Calibration


This accreditation is subject to with the above standard & IQAS requirements
The scope of accreditation is attached to the certificate & considered as part of it

The most recent issue of the accreditation scope is available on IQAS website
<https://iqas.mop.gov.iq>

Accreditation is valid From 10/7/2025 To 9/7/2027

Initial accreditation date
10/7/2025


Eng. Abdul Wahid M. Ibrahim
Director General of IQAS


Dr. Mohammed Ali Tamim
Deputy Prime Minister
Minister of Planning

	استمارة مجال الاعتماد Scope of Accreditation form	نظام الاعتماد العراقي IQAS
Organization address: Iraq - Kurdistan Region - Erbil	Organization name: Professional Laboratory Company for General Contracting Ltd	Accreditation no.: CL 016
Signature:  Eng. Abdul Wahid M. Ibrahim Director General of IQAS	Accreditation is valid: From 10/7/2025 To 9/7/2027	Issue no.: 001

Calibration Field	Type of Instrument or Material to be Calibrated or Measured	Measured or reference material	Calibration Method	Measurement Range	Uncertainty ($\pm$)	Location
Force	compression force load cell	Load cell 100 KN	DAKKS-DKD-R3-3	0– 100 KN	0.24	Lab and costumer premises
Force	compression force load cell	Load cell 3000 KN	DAKKS-DKD-R3-3	0– 3000 KN	0.24	Lab and costumer premises
Mass	Balance	Set of weights from 1 mg-10 kg	OIML R111-1 2004 (SOP 3-5-01)	1-10 kg	Class M 1	Lab and costumer premises
Temperature	Oven and water bath	k-type thermo couple	UMS SOP3 3-7-09	0-200 °C	0.1	Lab and costumer premises