

شهادة اعتماد

رقم TL 021



يقر نظام الاعتماد العراقي بأن:
مختبر الشركة الحديثة لخدمات فحص المنتجات
وتقييم ادارة الجودة

العراق - اقليم كردستان - اربيل - شارع كزنا الرئيسي

تم اعتماده وفقا لمتطلبات المواصفة ISO/ IEC 17025:2017
(المتطلبات العامة لاهلية مختبرات الفحص والمعايرة)

في مجال:
- اختبارات المواد المعدنية
- اختبارات مفاتيح الضغط

شرط التوافق مع متطلبات المواصفة اعلاه ومتطلبات IQAS الخاصة بالاعتماد
مجال الاعتماد المرفق بالشهادة يعتبر جزءا لا يتجزأ منها

يمكن الحصول على الاصدار الاحدث من مجال الاعتماد من خلال الموقع الالكتروني
<https://iqas.mop.gov.iq>

يكون الاعتماد نافذا من 2025/7/2 الى 2027/7/1
تاريخ منح الاعتماد لأول مرة
2018/4/16

لقد
أ.د. محمد علي تميم
نائب رئيس مجلس الوزراء
وزير التخطيط

المهندس
عبد الواحد محمد ابراهيم
مدير عام الهيئة/ وكالة

Ministry of planning
Iraqi Organization for Accreditation
IQAS

ACCREDITATION CERTIFICATE

No. TL 021



Iraqi Accreditation System Certify that:
**Modern Surveying Calibration
& Testing Labs**

Iraq - Kurdistan Region - Erbil - Main Gazna Road

Is accredited according to the requirements of the standard ISO/IEC 17025:2017
(General Requirements for the Competence of Testing and Calibration Laboratories)

In the field of:

- Metallurgical Materials Testing
- Pressure Switch's Testing

This accreditation is subject to with the above standard & IQAS requirements
The scope of accreditation is attached to the certificate & considered as part of it

The most recent issue of the accreditation scope is available on IQAS website
<https://iqas.mop.gov.iq>

Accreditation is valid From 2/7/2025 To 1/7/2027

Initial accreditation date
16/4/2018


Eng. Abdul Wahid M. Ibrahim
Director General of IQAS


Dr. Mohammed Ali Tamim
Deputy Prime Minister
Minister of Planning

	استمارة مجال الاعتماد Scope of Accreditation form	نظام الاعتماد العراقي IQAS
Organization address: Iraq - Kurdistan Region - Erbil - Main Gazna Road	Organization name: Modern Surveying Calibration & Testing Labs	Accreditation no.: TL 021
Signature: Eng. Abdul Wahid M. Ibrahim Director General of IQAS	Accreditation is valid: From 2/7/2025 To 1/7/2027	Issue no.: 005

Testing field	Type of test	Test object or product	Reference to standardized method
Mechanical	Determination of the seat tightness of metal and soft-seated pressure safety valves & pressure relief valves	Pressure Safety Valves & Pressure Relief Valves (PSV, PRV, PVRV and Pilot Valve etc.): -Pneumatic (-0.9 to 300 Bar) Hydraulic (0.2 to 1200 Bar)	MSL/TP/P/02 based on API 576 & API 527
Mechanical	On-line testing of pressure safety valves & relief valves for the determination of valve set pressure	Pressure Safety Valves & Pressure Relief Valves (PSV, PRV etc.): (1 Bar to 400 Bar)	MSL/TP/P/04 based on API 576 & API 527
Electro-Mechanical	To detect and correct errors in the set point and dead band of the pressure switch	Pressure Switches (Pneumatic) (-0.9 to 200 Bar) (Hydraulic) (0 to 1200 Bar)	MSL/TP/P/03