

شهادة اعتماد

رقم 017 CL



يقر نظام الاعتماد العراقي بأن:
مختبر شركة اليوناني للتجارة العامة المحدودة
العراق - اقليم كردستان - أربيل - الصناعة الجنوبية - طريق مخمور

تم اعتماده وفقا لمتطلبات المواصفة ISO/ IEC 17025:2017
(المتطلبات العامة لاهلية مختبرات الفحص والمعايرة)

في مجال:
- معايرة الجريان

شرط التوافق مع متطلبات المواصفة اعلاه ومتطلبات IQAS الخاصة بالاعتماد
مجال الاعتماد المرفق بالشهادة يعتبر جزءا لا يتجزأ منها

يمكن الحصول على الاصدار الاحدث من مجال الاعتماد من خلال الموقع الالكتروني

<https://iqas.mop.gov.iq>

يكون الاعتماد نافذا من 2025/10/20 الى 2027/10/19
تاريخ منح الاعتماد لأول مرة
2025/10/20

لقت

أ.د. محمد علي تميم
نائب رئيس مجلس الوزراء
وزير التخطيط

المهندس
عبد الواحد محمد ابراهيم
مدير عام الهيئة/ وكالة

Ministry of planning
Iraqi Organization for Accreditation
IQAS

ACCREDITATION CERTIFICATE

No. CL 017



Iraqi Accreditation System Certify that:

Laboratory of Al-Yonani Ltd Company

Iraq- Kurdistan region - Erbil – South Industry – Makhmour road

Is accredited according to the requirements of the standard ISO/IEC17025:2017
(General Requirements for the Competence of Testing and Calibration Laboratories)

In the field of:

- Flow Calibration

This accreditation is subject to with the above standard & IQAS requirements
The scope of accreditation is attached to the certificate & considered as part of it


The most recent issue of the accreditation scope is available on IQAS website


<https://iqas.mop.gov.iq>

Accreditation is valid From 20/10/2025 To 19/10/2027

Initial accreditation date

20/10/2025


Eng. Abdul Wahid M. Ibrahim
Director General of IQAS


Dr. Mohammed Ali Tamim
Deputy Prime Minister
Minister of Planning

	استمارة مجال الاعتماد Scope of Accreditation form	نظام الاعتماد العراقي IQAS
Organization address: Iraq- Kurdistan region - Erbil – South Industry – Makhmour road	Organization name: Laboratory of Al-Yonani Ltd Company	Accreditation no.: CL 017
Signature:  Eng. Abdul Wahid M. Ibrahim Director General of IQAS	Accreditation is valid: From 20/10/2025 To 19/10/2027	Issue no.: 001

Calibration Field	Type of Instrument or Material to be Calibrated or Measured	Measured or reference material	Calibration Method	Measurement Range	Uncertainty (±)	Location
Flow	Mass flow meter	INST.NAME: MASS FLOW METER MAKE/MODE L: Endress Hauser/Promass Q SERIAL NO. : X4016402000	API MPMS CHAPTER 4.8&5.6 CH11.1:2004	0 to 200000 KG/Hr	0.1%	Laboratory