

شهادة اعتماد

رقم 019 CL



يقر نظام الاعتماد العراقي بأن:
مختبرات قسم قياسات نقل الملكية/ شركة نفط الشمال
العراق- كركوك - عرفة - منطقة بابا الصناعية

تم اعتمادها وفقا لمتطلبات المواصفة ISO/ IEC 17025:2017
(المتطلبات العامة لاهلية مختبرات الفحص والمعايرة)

- في مجال:
- معايرة درجة الحرارة
 - معايرة الضغط
 - معايرة الجريان

شرط التوافق مع متطلبات المواصفة اعلاه ومتطلبات IQAS الخاصة بالاعتماد
مجال الاعتماد المرفق بالشهادة يعتبر جزءا لا يتجزأ منها

يمكن الحصول على الاصدار الاحدث من مجال الاعتماد من خلال الموقع الالكتروني
<https://iqas.mop.gov.iq>

يكون الاعتماد نافذا من 2025/12/17 الى 2027/12/16
تاريخ منح الاعتماد لأول مرة
2025/12/17

نفذ
أ.د. محمد علي تميم
نائب رئيس مجلس الوزراء
وزير التخطيط

محمد
محمد آيتن عمر
مدير عام الهيئة/ وكالة

Ministry of planning
Iraqi Organization for Accreditation
IQAS

ACCREDITATION CERTIFICATE

No. CL 019



Iraqi Accreditation System Certify that:
Laboratories of Custody Transfer Measurement
Dept./ North Oil Company
Iraq – Kirkuk - Arafa - Baba Industrial Zone

Is accredited according to the requirements of the standard ISO/IEC17025:2017
(General Requirements for the Competence of Testing and Calibration Laboratories)

In the field of:


- Temperature Calibration
- Pressure Calibration
- Flow Calibration


This accreditation is subject to with the above standard & IQAS requirements
The scope of accreditation is attached to the certificate & considered as part of it


The most recent issue of the accreditation scope is available on IQAS website
<https://iqas.mop.gov.iq>

Accreditation is valid From 17/12/2025 To 16/12/2027

Initial accreditation date
17/12/2025


Mohammed Ayden Omar
Director General of IQAS


Dr. Mohammed Ali Tamim
Deputy Prime Minister
Minister of Planning

	<p>استمارة مجال الاعتماد</p> <p>Scope of Accreditation form</p>	<p>نظام الاعتماد العراقي</p> <p>IQAS</p>
<p>Organization address: Iraq – Kirkuk - Arafa - Baba Industrial Zone</p>	<p>Organization name: Laboratories of Custody Transfer Measurement Dept./ North Oil Company</p>	<p>Accreditation no.: CL 019</p>
<p>Signature:  Mohammed Ayden Omar Director General of IQAS</p>	<p>Accreditation is valid: From 17/12/2025 To 16/12/2027</p>	<p>Issue no.: 001</p>

Calibration Field	Type of Instrument or Material to be Calibrated or measured	Measurand or reference material	Calibration method	Measurement range	Uncertainty (±)	Location
Temperature	Temperature Transmitter RTD Temperature Gauge	Dry Metrology Well	API/MPMS DKD INC	(-25) - (+150)C°	0.05	laboratory
Pressure	Pressure gauge Pressure Transmitter	Dead wight tester	DKD INC	0-300 bar	0.1	laboratory
Flow	Coriolis meters Pd meters	Portable master meter or master PD meter	API/MPMS INC	0-450 m ³ /hr	0.25%	NOC At site