وزارة التخطيط
الهيأة العراقية للاعتماد
IQAS

شـــهـادة اعتـــمـاد

رقم CL 002

يقر نظام الاعتماد العراقي بأن:

مختبر شعبة المعايرة/ قسم الفحص الهندسي والمختبرات /الشركة العامة للفحص والتأهيل الهندسي

العراق – بغداد- السيدية- طريق الحلة

تم اعتماده وفقا لمتطلبات المواصفة ISO/ IEC 17025:2017
(المتطلبات العامة لاهلية مختبرات الفحص والمعايرة)

في مجال:

- معايرة الابعاد
- معايرة الكتلة

شرط التوافق مع متطلبات المواصفة اعلاه ومتطلبات IQAS الخاصة بالاعتماد
مجال الاعتماد المرفق بالشهادة يعتبر جزءا لايتجزء منها

يمكن الحصول على الاصدار الاحدث من مجال الاعتماد من خلال الموقع الالكتروني
https://iqas.mop.gov.iq

يكون الاعتماد نافذا من ٢٠٢٦/٧/٢٢ الى ٢٠٢٨/٧/٢١
تاريخ منح الاعتماد لاول مرة
٢٠١٦/٣/٢٧

محمد ايدن عمر
مدير عام الهيأة/ وكالة

فالح ساري
وزير المالية
وزير التخطيط/ وكالة

Ministry of planning
Iraqi Organization for Accreditation
IQAS

ACCREDITATION CERTIFICATE

No. CL 002

Iraqi Accreditation System Certify that:

Laboratory of Calibration Section/ Engineering Inspection & Labs Department/ State Company for Inspection & Engineering Rehabilitation (SIER)

Iraq – Baghdad- Al-Syadia – Hilla road

Is accredited according to the requirements of the standard ISO/IEC 17025:2017 (General Requirements for the Competence of Testing and Calibration Laboratories)

In the field of:

- Dimensional Calibration

- Mass Calibration

This accreditation is subject to with the above standard & IQAS requirements
The scope of accreditation is attached to the certificate & considered as part of it
The most recent issue of the accreditation scope is available on IQAS website
https://iqas.mop.gov.iq

Accreditation is valid From 22/7/2026 To 21/7/2028
Initial accreditation date
27/3/2016

Mohammed Ayden Omar
Director General of IQAS

Falih Sari
Minister of Finance
Minister of Planning

	استمارة مجال الاعتماد Scope of Accreditation form	نظام الاعتماد العراقي IQAS
Organization address: **Iraq – Baghdad- Al-Syadia – Hilla road**	**Organization name:** **Laboratory of Calibration Section/ Engineering Inspection & Labs Department/ State Company for Inspection & Engineering Rehabilitation (SIER)**	**Accreditation no.:** **CL 002**
Signature: **Abdul Wahid M. Ibrahim** **Deputy General Manager**	**Accreditation is valid:** **From 22/7/2026 To 21/7/2028**	**Issue no.:** **006**

Calibration Field	Type of Instrument or Material to be Calibrated or Measured	Measure and or reference material	Calibration Method	Measurement Range	Uncertainty (±)	Location
Dimension	**Sieves**	**Profile projector** **Digital venire caliper**	**ISO 3310-1:2016** **ISO 3310-2:2013**	**20µm-≤5mm** **>5mm-≤125mm**	**± 0.024 mm** **± 0.038 mm**	**SIER laboratories**
Mass	**Weighing instrument Range >200g to 30kg**	**مجموعة اوزان دقة E2 من1mg-10kg** **وزن مفرد دقة E2 (20kg)** **مجموعة اوزان دقة F1 من1mg-1kg** **مجموعة اوزان دقةF1 من1mg-10kg** **وزن مفرد دقة F1 (20kg)** **مجموعة اوزان دقة F2 من1mg-10kg** **وزن مفرد دقة F2 (20kg)** **مجموعة اوزان دقة M1 من1mg-10kg** **وزن مفرد دقة M1 (20kg)**	**EA-cg18:2016** **OIMLR-76-1:2006**	**Cap.≤200g** **Cap.>200≤3kg** **Cap.>3kg≤6kg** **Cap.>6kg≤10kg** **Cap>10kg-≤30kg**	**± 0.1 mg** **± 1.5 mg** **± 5 mg** **± 10 mg** **± 20 mg**	**Costumer Premises**